IEC IECEE CB SCHEME

Ref. Certif. No.

DE 3 - 40682

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC

Product Produit	Treadmill-Ergometer
Name and address of the applicant Nom et adresse du demandeur	h/p/cosmos sports & medical gmbh Am Sportplatz 8 83365 Nussdorf - Traunstein, GERMANY
Name and address of the manufacturer Nom et adresse du fabricant	h/p/cosmos sports & medical gmbh Am Sportplatz 8, 83365 Nussdorf - Traunstein, GERMANY
Name and address of the factory Nom et adresse de l'usine	h/p/cosmos sports & medical gmbh Am Sportplatz 8, 83365 Nussdorf - Traunstein, GERMANY
Ratings and principal characteristics Valeurs nominales et caractéristiques principales	Rated Voltage: 200-240 V AC Rated Frequency: 50-60 Hz Rated Current: 6 A (Long time)/ 13,5 A (momentary) Protection Class: I Applied Part: B
Trade mark (if any) Marque de fabrique (si elle existe)	pluto®
Model/type Ref. Ref. de type	pluto® med, pluto® lt med
Additional information (if necessary) Information complémentaire (si nécessaire)	Excluded clause 17 (EMC); Certificate DE 3 - 40656 issued 2016-12-01 is replaced by this version due to non-technical changes.
A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012
as shown in the Test Report Ref. No. which forms part of this certificate comme indiqué dans le Rapport d'essais numéro de référence qui constitue une partie de ce certificat	713060696

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2017-01-20
CB 17 01 45283 019

Christoph Rappl

TÜV SÜD Product Service

TÜV SÜD Product Service GmbH · Certification Body · Ridlerstrasse 65 · D-80339 München

CB-1 03.06