



America

CERTIFICATE


No. U8 17 01 45283 020

Holder of Certificate: h/p/cosmos
sports & medical gmbh

Am Sportplatz 8
83365 Nussdorf - Traunstein
GERMANY

Production Facility(ies): 45283

Certification Mark:



Product: Treadmill-Ergometer

Model(s): pluto® med, pluto® lt med

Parameters:	Rated Voltage	200-240 V AC
	Rated Frequency	50-60 Hz
	Rated Current	6 A (longtime)/ 13,5 A (momentary)
	Protection Class	I
	Applied Part	B

Tested according to: ANSI/AAMI ES60601-1:2005/(R)2012
CAN/CSA-C22.2 No. 60601-1:14

The product was voluntarily tested according to the relevant safety requirements noted above. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited certification body.

Test report no.: 713060696

Date, 2017-01-20

Page 1 of 1

