



CERTIFICATE

No. U8 045283 0023 Rev. 00

Holder of Certificate: h/p/cosmos

sports & medical gmbh

Am Sportplatz 8

83365 Nussdorf - Traunstein

GERMANY

Production Facility(ies):

Certification Mark:

045283



Product:

Treadmill-Ergometer

Brand Name:

h/p/cosmos

Model(s):

pluto® med, pluto® It med, mercury med,

motion sprint 600 PL med

Parameters:

Rated voltage:

200 V – 240 V_{AC} or 120 V_{AC}

Rated frequency:

50 - 60 Hz

Rated electricity:

long time 6,0 A & momentary 13,5 A

long time 9.5 A & momentary 23.5 A

Protection class:

Type of applied part:

В

Software version:

MCU 5, v1.08.02

Tested according to: CAN/CSA-C22.2 No. 60601-1:14 ANSI/AAMI ES60601-1:2005/(R)2012

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.:

713141002

Date,

2019-07-02

(Paul Bielawski)