




America

CERTIFICATE

No. U8 045283 0025 Rev. 00

Holder of Certificate: h/p/cosmos
sports & medical gmbh
Am Sportplatz 8
83365 Nussdorf - Traunstein
GERMANY

Certification Mark:



Product: Treadmill-Ergometer

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing and Certification Regulations". TÜV SÜD America Inc. is an OSHA recognized NRTL and a Standards Council of Canada accredited Certification body.

Test report no.: 713184568

Date, 2022-01-31

(Paul Bielawski)



America

CERTIFICATE

No. U8 045283 0025 Rev. 00

Model(s):

- pulsar[®] med
- pulsar[®] It med
- stellar[®] 190/65 med
- pulsar[®] med OEM
- quasar[®] med
- quasar[®] It med
- stellar[®] med
- quasar[®] med OEM

Brand Name: h/p/cosmos

Tested according to: CAN/CSA-C22.2 No. 60601-1:2014
ANSI/AAMI ES60601-1:2005/(R)2012

Parameters:

| | |
|-----------------------------|--|
| Rated Voltage: | 200 V 2~, 208 V 2~, 230 V~ |
| Rated Frequency: | 50 – 60 Hz |
| Current Consumption: | (200 V 2~) = long time 10.5 A, momentary 23.0 A (208 V 2~) = long time 10.0 A, momentary 20.0 A (230 V~) = long time 7.0 A, momentary 17.0 A |
| Protection class: | I |
| Applied Part: | Type B |
| Software Version / Firmware | MCU 6, 1.2.0146 |