

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Treadmill-Ergometer
Name and address of the applicant	h/p/cosmos sports & medical gmbh Am Sportplatz 8 83365 Nussdorf - Traunstein GERMANY
Name and address of the manufacturer	h/p/cosmos sports & medical gmbh Am Sportplatz 8, 83365 Nussdorf - Traunstein, GERMANY
Name and address of the factory	h/p/cosmos sports & medical gmbh Am Sportplatz 8, 83365 Nussdorf - Traunstein, GERMANY
Ratings and principal characteristics	see page 2
Trade mark (if any)	h/p/cosmos
Model/type Ref.	pulsar [®] med pulsar [®] lt med stellar [®] 190/65 med pulsar [®] med OEM quasar [®] med quasar [®] lt med stellar [®] med quasar [®] med OEM
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005 IEC 60601-1:2005/AMD1:2012 IEC 60601-1-6:2010 IEC 60601-1-6:2010/AMD1:2013
as shown in the Test Report Ref. No. which forms part of this certificate	713184568

Page 1 of 2

This CB Test Certificate is issued by the National Certification Body

CB 045283 0026 Rev. 00

Date, 2022-02-01



(Paul Bielawski)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

Ratings and principal characteristics (continued)

Rated Voltage:	200 V 2~, 208 V 2~, 230 V~
Rated Frequency:	50 – 60 Hz
Current Consumption:	(200 V 2~) = long time 10.5 A, momentary 23.0 A (208 V 2~) = long time 10.0 A, momentary 20.0 A (230 V~) = long time 7.0 A, momentary 17.0 A
Protection class:	I
Applied Part:	Type B

Page 2 of 2

This CB Test Certificate is issued by the National Certification Body

CB 045283 0026 Rev. 00

Date, 2022-02-01



(Paul Bielawski)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany



Product Service