


America

CERTIFICATE

No. U8 045283 0029 Rev. 00

Holder of Certificate: h/p/cosmos
sports & medical gmbh
 Am Sportplatz 8
 83365 Nussdorf - Traunstein
 GERMANY

Certification Mark:




Product: Treadmill-Ergometer

This product was voluntarily tested to the relevant safety requirements referenced on this certificate. It can be marked with the certification mark above. The mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles system 3 as defined in ISO/IEC 17067. Certification is based on the TÜV SÜD "Testing, Certification, Validation and Verification Regulations (TCVVR)". TÜV SÜD America Inc. is an OSHA recognized NRTL for USA and a Standards Council of Canada ISO/IEC 17065 accredited Certification body for Canada.

Test report no.: 713309570

Date, 2024-06-06

(Paul Bielawski)



America

CERTIFICATE

No. U8 045283 0029 Rev. 00

Model(s):

- pluto® It med
- pluto® It med OEM
- pluto® ef med
- pluto® ef med OEM
- pluto® med
- pluto® med OEM
- stratos® med
- stratos® med OEM
- mercury® med
- mercury® med OEM

Brand Name: h/p/cosmos

Brand:



Tested according to:

CSA C22.2 No. 60601-1:2014/A2:2022-03
 CSA C22.2 No. 60601-1-6:2011/A2:2021-08
 ANSI/AAMI ES60601-1:2005/A2:2021

Parameters:

Electrical Ratings: 200 V - 240 V~, 50 Hz – 60 Hz
 (200-240 V~) current input: long time 6.0 A / momentary 13.5 A
 (100 V~) current input: long time 11.0 A / momentary 21.0 A
 (110 V~) current input: long time 11.0 A / momentary 18.0 A
 (120 V~) current input: long time 10.0 A / momentary 17.0 A
 (127 V~) current input: long time 10.0 A / momentary 16.0 A

Ingress Protection: IP20
Applied Part Type: B
Safe working load: 300 kg / 600 lbs