

EC Declaration of Conformity

assigned to devices of medical and
rehabilitation equipment

CE 0123

manufacturer: h/p/cosmos sports & medical gmbh

Am Sportplatz 8, DE 83365 Nussdorf-Traunstein/Germany, commercial register # HRB 7563 Traunstein
phone +49 86 69 86 42 0 fax +49 86 69 86 42 49 email@h-p-cosmos.com www.h-p-cosmos.com

notified body: TÜV SÜD Product Service GmbH, Ridlerstrasse 65, 80339 München / Germany, www.tuev-sued.de

product: Treadmill Ergometer (running machine)

Classification MDD Directive 93/42/EEC	Class IIb
Classification according to ISO 20957-1	S, I, A
UMDNS Code	14-141 Running Machine
GMDN Codes	33015 EXERCISER, TREADMILL, LINE-POWERED 36679 ERGOMETER, TREADMILL
NBOG Code	MD 1108 active rehabilitation devices

product family: treadmill h/p/cosmos 150/50 (mercury)

model name	article #	model name	article #
h/p/cosmos stratos lt med	cos30000va05	h/p/cosmos stratos med	cos30000va06
h/p/cosmos mercury lt med	cos30000va07	h/p/cosmos mercury med	cos30000va08
h/p/cosmos locomotion 150/50 E med	cos30001va01	h/p/cosmos gaitway II F	cos30002va01
h/p/cosmos locomotion 150/50 DE med	cos30001-01va02	h/p/cosmos gaitway II S	cos30002va02

We herewith declare under our sole responsibility that the above models / types meet the essential requirements of **Annex I** of the **European Council Directive 93/42/EEC**

The **CE** 0123 - mark gets affixed to the products.

This declaration of conformity is valid for all above mentioned devices which have been produced by h/p/cosmos on 01.02.2015 or after 01.02.2015. The validity of this declaration of conformity expires with the release of a new declaration of conformity if this is required due to technical amendments or due to legal amendments of the norms and standards – however, at the latest on the expiry date of the EC-certificate according to medical device directive 93/42/EEC with the certificate number G1 14 11 45283 014 on 31.01.2019.

DE 83365 Nussdorf-Traunstein, February 01, 2015

signed for and on behalf of h/p/cosmos sports & medical gmbh



Franz Harrer
President & CEO / R & D Manager



Michael Hesper
Safety officer for medical devices
German Medical Devices Act §30